

Certificate



Certificate no.

TA 50331502 01

License Holder:

Powerbox AB
P O box 148
SE-Gnesta 64622
Sweden

Manufacturing Plant:

ZTW1-LinPr-10052374 001

Test report no.: ZTW1-LinPr 10052374 001 **Client Reference:** WZ-16122015-OBS010xC

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product:	(Switching Power Supply)	License Fee - Units
Type Designation	: OBS010xC (POWERBOX)	5
x stands for	: 46, 50, 55, 60, 61, 64, 70	1
Rated Voltage	: AC 100-240V, 47-63Hz	
Rated Current	: 12.5-5.2A	
Output Voltage	: see appendix	
Output Current	: see appendix	
max. Ambient Temperature:	50°C for DC output full load 70°C for DC output de-rating load	
Protection Class	: I	

This certificate is valid for maximum 5 years.

6

Licensed Test mark:



TÜV Rheinland Taiwan Ltd

Signature



Dipl.-Ing. (FH) A. Klinker

Date of Issue

(day/mo/yr)

29/01/2016